

5		P	CALIB	RATI	ON	CERT	IFICAT	<u>E</u>		
CERTIFICATE N	0:	SBS/CL	/23/01726			ME	CHANICAL		Page No:1 of 1	
Issue Date					13-02	-2023		•		
SRF No & Date					SRF/23/00054-0001 & 10-02-2023					
Receipt Date					10-02-2023					
Calibration Date					10-02-2023					
Calibration Due					09-02-2024					
Customer Name	& Addr	ess								
GOVERNMENT	PRIMAR	Y HEALT	H CENTRE,							
KOTTAIYUR-630	106,SIV	AGANGA	DISTRICT.							
			Detail	ls of Devi	e Ur	der Calibrat	ion (DUC)			
Description	:	CENTR	FUGE		Make	& Model	:	KEMI & C8C		
Range	£ .	6000	RPM		Sr. N	0		12776		
_east Count	:	100	RPM		Ident	ification No	:			
DUC Condition	_ :	Satisfac	tory		Locat	tion		LAB		
		Envi	ronmental Co	nditions 8	Sta	ndard Opera	ting Procedu	re Details		
Environmental Details Temperature			e: 25.3 °C	: 25.3 °C Relative Hu		midity	45 % Rh			
Calibration Procedure No SBS/CP/ML			_/04	04 Calibration		one at ONSITE				
				Reference	e St	andards Deta	ails			
S.No Descriptio	No Description			Make/ SI No:		Certificate No		Validity		
1 Digital Tac	Digital Tachometer			LINE SEIKI / 175-0034V			TMS/22/63 01		10-11-2023	
CALIBRATION F	RESULTS	3								

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	2000	1998.6	1.4	10	
2	4000	3998.6	1.4	10	
3	6000	5998.5	1.5	10	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By (Calibration Engineer) S.MURALI



Quality Manager (D.VETRI SELVI)

Chief Executive

Authorised Signatory