



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2022



PC-1024

Lab Name Govt. Block PHC, T.Oranally Lab No 7542
 Constituent Group Chemistry III Date of Result Entered : 20/09/2022
 PT item Lyophilized Serum Date of Report Published : 05/10/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1846	223.49	6.67	14.90	216 mg/dL	-0.50	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	47	50.52	14.04	7.09	38.6 mg/dL	-1.68	2.07
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1834	3.49	9.05	0.32	3.5 mg/dL	0.03	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	112.08	5.53	6.20	115 mg/dL	0.47	0.29

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, T.Oranally

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NILGIRIS DISTRICT.

T.Oranally

TAMILNADU643003

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Signature]
 MEDICAL OFFICER,
 GOVT. PRIMARY HEALTH CENTRE,
 T. ORANALLY,
 THE NILGIRIS.