



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JANUARY 2023



PC-1024

Lab Name	Govt. Block PHC, T.Oranally	Lab No	7542
Constituent Group	Chemistry III	Date of Result Entered :	20/01/2023
PT item	Lyophilized Serum	Date of Report Published :	09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	125 mg/dL	0.67	0.35
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	172	41.93	10.86	4.55	41 mg/dL	-0.20	0.69
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	2.1 mg/dL	-0.50	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	113 mg/dL	0.13	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, T.Oranally
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NILGIRIS DISTRICT.
 T.Oranally
 TAMILNADU643003

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Signature]
 MEDICAL OFFICER,
 GOVT. PRIMARY HEALTH CENTRE,
 T. ORANALLY,
 THE NILGIRIS.