

CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	Calibration date : 11.01.2023
	Calibration valid: 10.07.2023
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s . GOVT. UPGRADED PRIMARY HEALTH CENTRE, T. ORANALLI-643003
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	
i Manufactured by	MICRO PIPETTE
ii Model	THERMO FISHER
iii Value	SCIENTIFIC
iv Serial No	10-100 µl
	RW13387

Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	08-OCT-2022	07-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)			
	10µl	50µl	100µl	100.14
1	10.03	50.06	100.14	100.14
2	10.04	50.09	100.12	100.12
3	10.05	50.02	100.10	100.10
4	10.09	50.08	100.08	100.08
5	10.00	49.94	100.09	100.09
6	10.08	50.07	100.12	100.12

For Heart Beat Medical Systems,

S. Anja



(Authorized signatory)