



CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	Calibration date : 11.01.2023
	Calibration valid: 10.07.2023
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s . GOVT. UPGRADED PRIMARY HEALTH CENTRE, T. ORANALLI-643003
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	FINNPIPETTE F3
II Model	100-1000 µl
III Serial No	RW13387

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	490.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

S. Aje

(Authorized signatory)

