



Abbott Healthcare Pvt. Ltd.  
Floor 13-19, Godrej BKC, Plot C-68  
G Block, Bandra Kurla Complex  
Near MCA Club, Bandra East  
Mumbai - 400051. India

Registered Office:  
4, Corporate Park, Sion  
Trombay Road, Mumbai  
- 400071. India

T : +91 22 3816 2000 / 1010  
F : +91 22 3816 2400 / 2401  
Email : webmaster@abbott.co.in  
Site : www.abbottdiagnostics.com  
CIN : U24200MH1997PTC104834

## **Preventive Maintenance & Calibration Certificate**

This is to inform that the Abbott Architect C4000 Fully Automated Biochemistry Analyser bearing the serial number C462207 Installed at Telangana Diagnostics HUB, Jangaon has been performed with the Preventive Maintenance and the Hardware Calibration of the instrument. The results are within the acceptable limit. The Preventive Maintenance and Hardware Calibration was performed on 22<sup>nd</sup> April 2022.

The next Preventive Maintenance and Hardware Calibration due is on 21<sup>st</sup> April 2023.

For Abbott Healthcare Pvt Ltd.

A handwritten signature in black ink, appearing to read 'P Nagendrakumar', on a light-colored background.

P Nagendrakumar  
(Technical Service specialist)

# Procedure Report

**Procedure:** 1120 Sample Pipettor Calibration

**Procedure version:** 12

**Module no.:** 1

**Serial no.:** c462207

**Operator ID:**

**Date:** 22.11.2022

**Time:** 15:14:52

**Status:** Completed

## Results:

Calibration Point	Theta	Z	Carousel	Carrier Position
Reaction Carousel	-5	32	2	NA
Wash Cup	-13	NA	NA	NA
Sample Handler	-11	19	NA	-6

Calibration passed.

# Procedure Report

**Procedure:** 1121 R1 Pipettor Calibration

**Procedure version:** 10

**Module no.:** 1

**Serial no.:** c462207

**Operator ID:**

**Date:** 22.11.2022

**Time:** 15:29:33

**Status:** Completed

## Results:

Calibration Point	Theta	Z	Carousel
Wash Cup	1	NA	NA
Reaction Carousel	0	NA	2
Outer Segment	0	8	34
Inner Segment	0	6	25

Calibration passed

# Procedure Report

**Procedure:** 1122 R2 Pipettor Calibration

**Procedure version:** 9

**Date:** 22.11.2022

**Module no.:** 1

**Time:** 15:43:57

**Serial no.:** c462207

**Status:** Completed

**Operator ID:**

## Results:

Calibration Point	Theta	Z	Carousel
Wash Cup	2	NA	NA
Reaction Carousel	1	NA	2
Outer Segment	1	5	32
Inner Segment	1	7	25

Calibration passed

# Procedure Report

**Procedure:** 2134 Change Water Bath

**Procedure version:** 6

**Module no.:** 1

**Serial no.:** c462207

**Operator ID:** FSE

**Date:** 22.11.2022

**Time:** 16:16:22

**Status:** Completed

## Results:

2134 Change Water Bath procedure completed.

# Procedure Report

**Procedure:** 3525 Temperature Status

**Procedure version:** 4

**Module no.:** 1

**Serial no.:** c462207

**Operator ID:** FSE

**Date:** 22.11.2022

**Time:** 16:07:04

**Status:** Completed

## Results:

**Processing Area**

**Temperature**

**Status**

Processing Area	Temperature	Status
Water Bath	36.90	OK
Reagent Supply Center.	5.80	OK
Instrument Internal.	28.20	OK

# Procedure Report

**Procedure:** 1114 Carrier Transport Calibration

**Procedure version:** 13

**Date:** 22.11.2022

**Module no.:** 0

**Time:** 16:06:39

**Serial no.:** RSH62207

**Status:** Completed

**Operator ID:** FSE

**Results:**

Position	Result
X at bay 1	2885
X at bay 2	7893
X at bay 3	12897
X at bay 4	17914
Theta at positioner	73
Z at positioner pocket 1	8047
X at positioner pocket 1	14101
Z at positioner pocket 2	8051
X at positioner pocket 2	9073
Z at bay 1	8017
Z at bay 2	8000
Z at bay 3	8009
Z at bay 4	8026