



Date: 10/01/2023

Certificate of Calibration

Customer Name: Ujala Cygnus Healthcare, Agra

Model: Fully Automated Biochemistry Analyzer EM-200

Serial No.: 2002715

Calibration done date: 10/01/2023

Next Calibration due Date: On or Before: 09/01/2024


Lab Incharge: Dr. Prachi Mehta

*This is to certify that above mentioned product has been verified of calibration
By*

*Checking Lamp Voltage -11.79
Checking gain of Photo meter for all Filters-OK
Checking Cell blanks value-OK
Checking laundry section-OK
Checking movement of probe-OK*

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of
QC*

Calibration on site performed by Pushpendra Singh Bhadauria


Pushpendra Singh Bhadauria
Sr. Service Engineer
Agra

