



YEA

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

☎ : 033 2677 1412, 2667 1909, 82329 85881

Tele Fax : 033 2677 1411

Admn. Office & Lab : 44, Shahid Khudiram Bose Sarani, (formerly Kamardanga Road), Ichapur, Howrah - 711 104

E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org

CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Calibration Certificate

Certificate No.: YEA(C)/ 230209A/I859/001
ULR No.: CC222123000012082F

Dated : 22.02.2023

Calibrated for : ICTC Khawzawl CHC
Khawzawl,
Pin - 796310.

Calibrated on.	Next Calibration Due On.	Page No.	No. of Pages
11.02.2023	11.02.2024	1	2

Description of Instrument

Name : Glass Thermometer
 Identification : KZ/ICTC/GT-01
 Range : -10 to 110 °C
 Least Count : 1 °C
 Make : N/F
 Model / Type : Alcohol

Service Request No. / Order Reference : 230209A/I859/001-002

Receipt of Instrument : 09.02.2023

Location where calibration performed : At permanent laboratory

Applicable Specification of Instrument to be Calibrated : Not Specified

Condition of Instrument to be Calibrated : Satisfactory

Basis of Calibration : Comparison method with reference to YEA/WI/T-01

Environmental Condition during Calibration : Temp. : 25±2°C RH : 50±15%

Equipment used for Calibration : Standard Temperature Calibrator & Standard RTD Sensor

Traceability : To National Standard vide Certificate No.
YEA(C)/230125/Y001/001 valid till 24.01.2024
TL/023/52.1.2 valid till 12.01.2024

Measurement Uncertainty : The figures are stated in the result at coverage factor
 $k=2$ corresponding to confidence probability of
95% for normal distribution.

Laboratory Identification : Sticker is affixed for authentication of calibration.

The above Instrument/Equipment/Machine has been calibrated.
The results are given in the page 2 to 2 ...of the Certificate.

2

YEA(C)/F-08



CC-2221



YEA

☎ : 033 2677 1412, 2667 1909, 82329 85881

Tele Fax : 033 2677 1411

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : 44, Shahid Khudiram Bose Sarani, (formerly Kamardanga Road), Ichapur, Howrah - 711 104

E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org

CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Certificate No.: YEA(C)/ 230209A/1859/001

Date: 22.02.2023

ULR No.: CC222123000012082F

Page 2 of 2 Pages

Calibrated By : M. Sadhukhan

Calibration Results

Thermal Calibration

Calibration Results of Glass Thermometer

Sl. No.	DUC Value	Standard Value	Measurement Uncertainty
	(°C)	(°C)	(± °C)
1	-10	-9.95	0.60
2	0	0.11	
3	49	49.19	
4	100	100.28	
5	110	110.37	

DUC : Device Under Calibration

REMARKS : The above Glass Thermometer has been calibrated and the readings are tabulated above.

NOTES : This certificate refers to the results tabulated at the time of calibration only for the above mentioned item submitted for calibration and this should not be reproduced, other than in full except with prior written approval from the head of calibration laboratory.

Authorised By

A. Das

Manager Calibration

*** End of the Calibration Certificate ***

