



YEA

☎ : 033 2677 1411/1412, 2667 1909

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : Kamardanga Road (Opp. Tarama Ashram), Ichapur, Howrah - 711 104

E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org

CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Calibration Certificate

Certificate No.: YEA(C)/ 230209A/I862/001
ULR No. : CC222123000012089F

Dated : 10.02.2023

Calibrated for : ICTC Sakawrdai CHC
Sakawrdai, Aizawl,
Pin - 796111.

Calibrated on.	Next Calibration Due On.	Page No.	No. of Pages
10.02.2023	10.02.2024	1	2

Description of Instrument

Name : Digital Thermometer with Sensor
Identification : SKD/ICTC/GT-01
Range : (-)50 to 300 °C
Least Count : 0.1 °C
Make : N / F

Service Request No. / Order Reference : 230209A/I862/001-002

Receipt of Instrument : 09.02.2023

Location where calibration performed : At permanent laboratory

Applicable Specification of Instrument to be Calibrated : Not Specified

Condition of Instrument to be Calibrated : Satisfactory

Basis of Calibration : Comparison method with reference to YEA/WI/T-02

Environmental Condition during Calibration : *Temp.* : 25±2°C *RH* : 50±15%

Equipment used for Calibration : Standard Temperature Calibrator & Standard RTD Sensor

Traceability : To National Standard vide Certificate No.
YEA(C)/230125/Y001/001 valid till 24.01.2024
TL/023/52.1.2 valid till 12.01.2024

Measurement Uncertainty : The figures are stated in the result at coverage factor
 $k= 2$ corresponding to confidence probability of
95% for normal distribution.

Laboratory Identification : Sticker is affixed for authentication of calibration.



CC-2221

The above Instrument/Equipment/Machine has been calibrated.
The results are given in the page 2 to 2 ...of the Certificate.

Cont.....2

YEA(C)/F-08





YEA

☎ : 033 2677 1411/1412, 2667 1909

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : Kamardanga Road (Opp. Tarama Ashram), Ichapur, Howrah - 711 104

E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org

CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Certificate No.: YEA(C)/ 230209A/1862/001

ULR No. : CC222123000012089F

Calibrated By : K. Rout

Date: 10.02.2023

Page 2 of 2 Pages

Calibration Results

Thermal Calibration

Calibration Results of Digital Thermometer with Sensor

Sl. No.	DUC Value	Standard Value	Measurement Uncertainty
	(°C)	(°C)	(± °C)
1	-23.2	-23.37	0.17
2	0.8	0.45	
3	99.3	98.61	
4	200.4	199.58	0.32
5	295.8	294.36	

DUC : Device Under Calibration

REMARKS : The above Digital Thermometer with Sensor has been calibrated and the readings are tabulated above.

NOTES : This certificate refers to the results tabulated at the time of calibration only for the above mentioned item submitted for calibration and this should not be reproduced, other than in full except with prior written approval from the head of calibration laboratory.

Authorised By

A. Das
Manager Calibration

*** End of the Calibration Certificate ***

