



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



PC-1024

Lab Name Govt. PHC, Nagavedu Lab No 13018
Constituent Group Chemistry III Date of Result Entered : 12/01/2023
PT Item Lyophilized Serum Date of Report Published : 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	C _v
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	114 mg/dL	-0.77	0.35
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	172	41.93	10.86	4.55	42 mg/dL	0.02	0.69
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	2.3 mg/dL	0.41	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	110 mg/dL	-0.32	0.32

SDI Range	Interpretation
thin -1.00 to +1.00	Excellent.
thin ± 1.01 to ± 2.00	Good.
thin ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
yond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Nagavedu
Nagavedu
MILNADU

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

மருத்துவ அலுவலர்,
அரசு ஆரம்ப சுகாதார நிலையம்
நாகவேடு-631 002
இராணிப்பேட்டை மாவட்டம்.