

Timer Calibration Verification Record Sheet

Location: ICTC Lab

Property No: MICO Hosp, Nasik (New Timer)

Procedure using Timer: Inhouse verification of timer using certified standard calibrated Timer

Date of physical verification: 19/07/2022

Due date of physical verification: 18/07/2023

Certified timer ID: CRV/MICRO/VIRO/03/2017-18, DT-02

Verification Done by: Ms Bandgar Rupali, Technical Officer, SRL, AFMC, Pune

Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)
1	300	300	0	600	601	1	900	901	1
2	300	300	0	600	602	2	900	900	0
3	300	300	0	600	600	0	900	901	1
Mean	300	300	0	600	601	1	900	900.66667	0.6666667
SD	-	0.00	-	-	1.00	-	-	0.58	-
%CV	-	0	-	-	0.1663894	-	-	0.0641025	-
%	-	0	-	-	0.1666667	-	-	0.0740741	-

Accept/ Reject: Accepted since average difference is within +/- 60 Sec

Final Interpretation: Timer is in good working condition

Action Taken: - NA

Remark: - NA

B. Bandgar
09/08/2023
Ms. Rupali Bandgar
Technical Officer
SRL, MSACS
AFMC, Pune - 411 040


ICTC
MEDICAL OFFICER
Mico Maternity Hospital
N.M.C. Satpur, Nashik