


Digital Thermometer Calibration Verification Record Sheet				
Location: ICTC lab				
Property No: MICO Satpur (New Thermometer)				
Procedure using Digital Thermometer: In house verification of Digital thermometer using standard calibrated thermometer				
Date of physical verification: 19/07/2022				
Due date of physical verification: 18/07/2023				
Certified thermometer ID: CRV/MICRO/VIRO/03/2017-18, 201705039234				
Verification Done by: Ms Bandgar Rupali, Technical Officer, SRL, AFMC, Pune				
Range of refrigerator: 2-8 Degree Celcius				
Sr.No.	Time	Test Temp Degree Celcius	Certified Temp Degree Celcius	Temp Difference Degree Celcius
1	11.00 am	6.1	5	-1.1
2	11.30 am	7.1	6.1	-1
3	12.00 pm	6.8	3.4	-3.4
4	12.30 pm	6.9	3.9	-3
Mean	-	6.725	4.6	-2.125
SD	-	-	1.20	-
%CV	-	-	26.15	-
% Accuracy=	-	-	-31.59851301	-
Accept/ Reject: Accepted since average temperature difference is between +/- 2.5 Degree Cencius				
Final Interpretation: Digital thermometer is in good working condition				
Action Taken: NA				
Correction factor:	-2.1			
Remark: NA				


 09/08/2023
 Ms. Rupali Bandgar
 Technical Officer
 SRL, MSACS
 AFMC, Pune - 411 040


 ICTC
 MEDICAL OFFICER
 Mico Maternity Hospital
 N.M.C. Satpur, Nashik