



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

ALL ANALYSER SUMMARY REPORT - AUGUST 2022



PC-1024

Lab Name **REDCLIFFE LIFETECH PVT LTD** Lab No **15558**
 Constituent Group **HbA1c** Date of Result Entered **17/08/2022**
 PT item **Lyophilized Whole blood** Date of Report Published **05/09/2022**

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	130	5.98	5.12	0.31			0.05
			Beckman AU480/680/5800/DXC700AU	32	6.13	5.37	0.33			0.12
			Any Analyser	260	6.58	11.29	0.74	6.9 %	0.43	0.09

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD

Mr. Ravi Shankar Dubey, SHREE TOWER, FLAT NO 102, DASARATHA, MAIN BY PASS ROAD, BEUR

PATNA

BIHAR,800002

Coordinator Contact details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****