

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/22/14964	MEDICAL DEVICES	Page No:1 of 1
Issue Date	26-12-2022	
SRF No & Date	SRF/22/00347 -0003 & 23-12-2022	
Receipt Date	23-12-2022	
Calibration Date	23-12-2022	
Calibration Due	22-12-2023	

Customer Name & Address

GOVERNMENT UPGRADED PRIMARY HEALTH CENTRE,
VILATHIKULAM BLOCK,KULATHUR-628903,THOOTHUKUDI DISTRICT.

Details of Device Under Calibration (DUC)

Description	ELECTRICAL SAFETY (CELL COUNTER)	Make & Model	SYSMEX & XP-100
Range		Sr. No	B5627
Resolution		Identification No	
DUC Condition	SATISFACTORY	Location	LAB

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 24.8° C	Relative Humidity	48% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	FLUKE & 2244202	ACCS221224	24-06-2023

RESULTS


Electrical Safety

S.no	Specification	Measured values in MΩ	Allowable limit in MΩ	Uncertainty in % (±)
1	Insulation Resistance	81	>20MΩ	13.92
		Measured values in µA	Allowable limit in µA	Uncertainty in % (±)
2	Earth Leakage	162	<5000µAfor B,BF,CF	5.9
		Measured values in µA	Allowable limit in µA	Uncertainty in % (±)
3	Enclosure Leakage	197	<500µAfor B,BF,CF	6.9

REMARKS

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2 The user should determine the suitability of the instrument for its intended use.
- 3 The recalibration interval should be determined on the user requirement.
- 4 The results stated in this certificate relate only to the item calibrated.
- 5 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6 Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By


(Calibration Engineer)
C SIVABALAN



Authorised Signatory


Quality Manager
(D.VETRI SELVI)

Chief Executive