

VIEW LAB MONTHLY SUMMARY

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Lab Name Govt. PHC, Nagamangalam
 Lab No 12728
 Month January
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/01/2023

Date of Report Published : 09/02/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|----------------------------------------------|--------------------|--------|--------------|------|------------|------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1890 | 119.87 | 6.38 | 7.65 | 125 mg/dL | 0.67 | 0.35 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1865 | 2.21 | 10.02 | 0.22 | 2.38 mg/dL | 0.77 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1794 | 112.14 | 6.03 | 6.76 | 115 mg/dL | 0.42 | 0.32 |

| SDI Range | Interpretation |
|-----------------------|------------------------------------------|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|------------------------------------------|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |

Green
Medical Officer
 GOVT PRIMARY HEALTH CENTRE
 NAGAMANGALAM
 TRICHY 620 017