

CALIBRATION CERTIFICATE

Name of the customer	INDIGLOBAL LABS PVT LTD PLOT NO 7,CF AREA,PHASE-II, IDA,CHERALAPALLY,HYDERABAD-500051 TELANGANA
----------------------	--

Make:	HORIBA MEDICAL FRANCE
Product Name:	ABX PENTRA C 200
S.No.	1911C2- 0209-D

This is to certify that the ABX PENTRA C200 Fully automated clinical Biochemistry analyzer has been duly inspected and calibrated on 11-06-2022. The instrument is working satisfactorily and the next calibration due on 10-06-2023.

S.No.	Test Parameters	Measured	Remarks			
1	IRU -1 Temp. (37.0 +/- 0.5 Deg C.)	37.09	Pass			
2	IRU -2 Temp. (37.0 +/- 0.5 Deg C.)	36.84	Pass			
3	SRCU Temp. (8.0-15.0 Deg C.)	8.106	Pass			
4	Software Version	1002	Pass			
5	Precision test (CV<=2%) Appendix 1	0.88	Pass			
6	Spectro gains (2to10) Appendix 2					
	Wavelength					
	a)340nm	9.010	Pass	e)570nm	8.880	Pass
	b)405nm	9.170	Pass	f)600nm	8.960	Pass
	c)510nm	9.030	Pass	g)660nm	8.910	Pass
	d)546nm	9.030	Pass	h)700nm	8.890	Pass

For HORIBA INDIA PVT LTD


M SURENDRA REDDY
(ENGINEER-CUSTOMER SUPPORT)