

Date: 11/06/2022

CALIBRATION CERTIFICATE

This is to state that the haematology cell counter model **HORIBA MEDICAL: YUMIZEN H500**, bearing serial no: **201YOXH04453** installed at **Indiglobal Labs Pvt Ltd, Hyderabad** was calibrated with Calibrator on **11th June 2022**.

Subsequently controls were processed and found them in respective range.

Calibrator Used: **ABX MINOCAL**

Lot No.: **CX471**

Expiry Date: **05/07/2022**

Parameter	Target Value (As per Kit Insert)	Mean Value	Observed CV%	Acceptance CV%	Comments
WBC	8.40	8.29	2.18	<2.5%	PASSED
RBC	4.60	4.57	1.51	<2%	PASSED
HGB	13.2	12.9	0.24	<1.5%	PASSED
HCT	40.5	40.2	1.52	<2%	PASSED
PLT	251	250	2.34	<5%	PASSED

Next Calibration Due Date: 10/06/2023



For **HORIBA MEDICAL,**
HORIBA INDIA PVT. LIMITED.