



112th IAMM EQAS Microbiology: Bacteriology/ Serology
CMC MICRO EQAS
 Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu
 Email: egsas@cmcvellore.ac.in, Twitter: @microegas, Phone: +91-416-2282588



NOVEMBER 2022

112th EQAS EVALUATION REPORT

MEMBER ID:

M	1	3	1	6
---	---	---	---	---

Marks Obtained: 9/10 (90%)

NOVEMBER 2022 / SEROLOGY

Please refer the attached evaluation format/answer template for details on the criteria for evaluation.

Parameter	Your Interpretation		Intended Result				Max Marks		Your Score
	Correct	Incorrect	STO	STH	Interpretation	STO	STH		
SE1 Widal*	✓		Negative	Negative	Not Suggestive of Enteric Fever / Typhoid Fever	2	2	4	

*Negative at 1:20 dilution

Parameter	Your Result	Intended result	correct	Incorrect	Max Marks	Your Score
RPR/VDRL	Non-Reactive	Non-Reactive	✓		2	2
TPHA	Non-Reactive	Non-Reactive	✓		2	2
SE2 Syphilis ELISA	NOT DONE					

SE3: Test method employed for detection C-reactive protein (CRP) at your lab: Turbidimetry
Peer group (n) = 444

Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score
SE3 CRP	Positive	142.96	Positive	67.8083	11.2497	0.63 to 691.0	0.6674	6.7	2	1

Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M 1 3 1 6

SE1	SE2	SE3	Marks obtained	
4	4	1	9	90%
4	4	2	Maximum marks =10	



Dr. Rani Diana Sahni
Scientific Co-ordinator

Report Dispatch Date: 15.02.23



Dr. John A Jude Prakash
Quality Manager



Dr. V. Balaji
PT Co-ordinator

..... End of Report