



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name	<b>TELANGANA DIAGNOSTIC HUB</b>	Lab No	<b>15756</b>
Constituent Group	<b>HbA1c</b>	Date of Result Entered :	<b>17/12/2022</b>
PT Item	<b>Lyophilized human whole blood based</b>	Date of Report Published :	<b>05/01/2023</b>

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Enzymatic	Abbott	16	4.02	8.75	0.35	4.1 %	0.23	0.18

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :

**TELANGANA DIAGNOSTIC HUB**  
**DISTRICT HOSPITAL PREMISES**  
**ASIFABAD**  
**TELANGANA504293**

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

