

# HORIBA

Est. 1962

## HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2023-2024/100737

10<sup>th</sup> Feb 2023

## CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX YUMIZEN H550** bearing serial number: **110YAXH03533** installed at **P.H. Medical Centre, Santacruz** calibrated on **10<sup>th</sup> Feb 2023**


Calibrator : ABX MINOCAL

Lot No. : CX480

Expiry Date : 05<sup>th</sup> April 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **09<sup>th</sup> Feb 2024**.



**Shrish Dixit**

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**

## Blank Cycle Logs

Running Date	Operator	WBC 10 <sup>9</sup> /L	RBC 10 <sup>6</sup> /L	HGB g/dL	PLT 10 <sup>9</sup> /L	Status	Technical alarms
02/10/2023 10:40:14 AM	DIPALI	0.00	0.00	0.0	4	Passed	
02/10/2023 07:29:19 PM	technician	1.56 <b>H</b>	0.01	0.0	12 <b>H</b>	Failed	
02/10/2023 07:29:29 PM	technician	0.02	0.00	0.0	8 <b>H</b>	Failed	
02/10/2023 07:42:59 PM	technician	0.01	0.00	0.0	4	Passed	
02/10/2023 07:47:37 PM	technician	0.00	0.00	0.0	5	Passed	

Manufactured By:  
**HORIBA India Private Limited**  
 (A subsidiary of HORIBA Limited Japan)  
 Plot No.26 Sector-7, I.I.E., SIDCUL,  
 Haridwar-249403, Uttarakhand, India  
 Toll Free No.: 1800 103 4477

**ABX Minoccal**



**LOT** CX 400

**CAL**

**(EXP.) 2023-04-05**

PARAMETERS	UNITS	VARIABLES		TOLERANCES		
		HSO	HSO/OT	HSO/CT	HSO/CT	
WBC	$10^9/mm^3 \times 10^9/L$	8.70	8.70	8.60	8.80	$\pm 0.20$
RBC	$10^{12}/mm^3 \times 10^{12}/L$	4.57	4.57	4.53	4.53	$\pm 0.06$
HGB	g/dl	13.2	13.2	13.0	13.0	$\pm 0.2$
HCT	%	38.2	38.2	37.5	38.2	$\pm 1.0$
PLT	$10^9/mm^3 \times 10^9/L$	0.382	0.382	0.375	0.382	$\pm 0.010$
WBC REF	$10^9/L$	10.8	10.8	9.2	9.2	$\pm 0.5$

Blood Control Vial should be disposed off as per State Government Bio-Medical Waste Management Rule.

**HORIBA**

## Repeatability Report (part 1)

Number of repeatability run report selected for statistic calculation 10/11

Coefficient	WBC (10 <sup>9</sup> /μL)	RBC (10 <sup>9</sup> /μL)	HGB (g/dL)	HCT (%)	PLT (10 <sup>9</sup> /μL)	MCV (μm <sup>3</sup> )	RDW-CV (%)	RDW-SD (μm <sup>3</sup> )	P-LCR (%)
Minimum	6.24	4.67	14.2	40.9	289	87.5	12.3	44.3	36.2
Maximum	6.70	4.79	14.4	42.0	317	88.1	13.2	46.6	39.8
Mean	6.41	4.72	14.3	41.5	301	87.7	12.7	45.3	37.9
Difference	0.47	0.12	0.2	1.1	28	0.6	0.9	2.3	3.6
2 SD	0.28	0.07	0.1	0.7	20	0.4	0.5	1.4	2.0
CV(%)	2.20	0.79	0.38	0.82	3.34	0.22	2.02	1.55	2.61

Sel	Run Date &Time	WBC (10 <sup>9</sup> /μL)	RBC (10 <sup>9</sup> /μL)	HGB (g/dL)	HCT (%)	PLT (10 <sup>9</sup> /μL)	MCV (μm <sup>3</sup> )	RDW-CV (%)	RDW-SD (μm <sup>3</sup> )	P-LCR (%)	Operator
	02/10/2023 07:49:12 PM	6.43	4.82	14.3	42.4	305	87.9	12.8	45.9	39.0	ABX
✓	02/10/2023 07:50:43 PM	6.49	4.79	14.4	42.0	307	87.7	12.6	45.1	38.3	ABX
✓	02/10/2023 07:52:19 PM	6.29	4.70	14.3	41.4	291	88.1	13.2	46.6	38.7	ABX
✓	02/10/2023 07:53:51 PM	6.57	4.75	14.4	41.6	317	87.6	12.9	45.9	39.8	ABX
✓	02/10/2023 07:55:26 PM	6.39	4.72	14.3	41.5	292	87.9	12.8	45.9	36.2	ABX
✓	02/10/2023 07:57:01 PM	6.70 *	4.75	14.3	41.8	289	87.9	12.7	45.1	38.0	ABX
✓	02/10/2023 07:58:34 PM	6.33	4.67	14.3	40.9	310	87.6	12.4	44.3	37.9	ABX
✓	02/10/2023 08:00:06 PM	6.24	4.73	14.3	41.5	299	87.7	12.6	45.1	38.2	ABX
✓	02/10/2023 08:01:42 PM	6.45	4.74	14.2	41.4	310	87.5	12.3	44.3	37.9	ABX
✓	02/10/2023 08:03:14 PM	6.34	4.67	14.3	41.0	301	87.6	12.7	45.1	37.5	ABX
✓	02/10/2023 08:04:44 PM	6.35	4.72	14.3	41.4	290	87.7	12.7	45.1	36.8	ABX

## Repeatability Report (part 2)

Number of repeatability run report selected for statistic calculation 10/11

Coefficient	NEU% (%)	LYM% (%)	MON% (%)	EOS% (%)	BAS% (%)	LIC% (%)	
Minimum	60.6	23.1	4.2	7.4	0.5	0.5	
Maximum	62.9	26.4	5.1	8.5	1.1	0.9	
Mean	61.7	25.0	4.6	7.9	0.8	0.7	
Difference	2.3	3.3	0.9	1.1	0.6	0.4	
2 SD	1.9	2.0	0.5	0.9	0.3	0.3	
CV(%)	1.50	3.93	5.08	5.46	21.67	19.43	

Sel	Run Date &Time	NEU% (%)	LYM% (%)	MON% (%)	EOS% (%)	BAS% (%)	LIC% (%)	Operator
	02/10/2023 07:49:12 PM	61.6	25.5	4.6	7.3	1.0	0.9	ABX
✓	02/10/2023 07:50:43 PM	62.2	24.2	4.2	8.3	1.1	0.7	ABX
✓	02/10/2023 07:52:19 PM	60.6	26.4	4.7	7.4	0.9	0.6	ABX
✓	02/10/2023 07:53:51 PM	62.5	24.8	4.6	7.4	0.7	0.5	ABX
✓	02/10/2023 07:55:26 PM	62.7	24.5	4.4	7.7	0.7	0.9	ABX
✓	02/10/2023 07:57:01 PM	62.5	24.6	4.7	7.5	0.7	0.6	ABX
✓	02/10/2023 07:58:34 PM	61.1	26.0	4.6	7.7	0.6	0.7	ABX
✓	02/10/2023 08:00:06 PM	62.9	23.1	4.7	8.5	0.8	0.8	ABX
✓	02/10/2023 08:01:42 PM	61.1	25.2	5.1	7.8	0.8	0.7	ABX
✓	02/10/2023 08:03:14 PM	60.6	25.6	4.5	8.5	0.8	0.9	ABX
✓	02/10/2023 08:04:44 PM	60.9	25.8	4.7	8.1	0.5	0.9	ABX

## Calibration Report

Sample ID CX480  
Lot number CX480

Name ABX MINOCAL

Exp. date 04/05/2023  
Modified on:

Coefficient	WBC	RBC	HGB	HCT	PLT	MPV
New	1.118	0.936	0.988	1.023	1.012	1.075
Current	1.150	0.930	0.988	1.027	0.988	1.065
Target	8.70	4.57	13.2	38.2	250	10.8
Mean	8.95	4.54	13.2	38.1	244	10.7
CV(%)	1.37	0.52	0.28	0.42	3.56	1.62

Number of calibration run selected for coefficient calculation (minimum 5) 5/6

Sel.	Run Time	WBC ( $10^9/\mu\text{L}$ )	RBC ( $10^9/\mu\text{L}$ )	HGB (g/dL)	HCT (%)	PLT ( $10^9/\mu\text{L}$ )	MPV ( $\mu\text{m}^3$ )
✓	02/10/2023 08:08:31 PM	8.81	4.57	13.2	38.1	254	10.8
	02/10/2023 08:10:02 PM	<b>9.06</b> h	4.51	13.3	37.8	244	11.0
✓	02/10/2023 08:11:34 PM	<b>9.14</b> h	4.55	13.2	38.0	252	10.7
✓	02/10/2023 08:13:06 PM	<b>8.94</b> h	<b>4.51</b> l	13.2	38.0	242	10.7
✓	02/10/2023 08:14:37 PM	<b>8.97</b> h	4.52	13.2	38.0	<b>295</b> l	10.7
✓	02/10/2023 08:16:11 PM	8.90	4.54	13.1	38.4	<b>297</b> l	10.4

**QC - Control Run Report**

Run Date 02/10/2023 08:19:44 PM


Operator ABX

Name NORMAL CONTROL  
 Level Normal  
 Lot number PX439N


Sample ID PX439N  
 Exp. date 03/31/2023

			<b>Range</b>
RBC	4.66	10 <sup>6</sup> /μL	4.42 - 4.82
HGB	13.3	g/dL	12.9 - 13.9
HCT	40.4	%	38.2 - 42.2
MCV	86.7	μm <sup>3</sup>	82.0 - 92.0
MCH	28.6	pg	27.0 - 31.0
MCHC	33.0	g/dL	30.3 - 36.3
RDW-CV	12.5	%	10.0 - 18.0


**Alarms**



			<b>Range</b>
PLT	264	10 <sup>3</sup> /μL	226 - 286
MPV	9.4	μm <sup>3</sup>	7.1 - 11.1



			<b>Range</b>
WBC	8.54	10 <sup>3</sup> /μL	7.20 - 9.20
	<b>#</b>	<b>Range</b>	<b>%</b>
NEU	4.07	2.98 - 4.78	47.8
LYM	3.70	2.85 - 4.25	43.3
MON	0.38	0.00 - 0.86	4.5
EOS	0.34	0.00 - 0.50	3.9
BAS	0.05	0.00 - 0.18	0.5



**Slide Review**

Neutrophil	Myeloblast	Anisocytosis
Lymphocyte	Promyelocyte	Hypochromia
Monocyte	Myelocyte	Polychromasia
Eosinophil	Metamyelocyte	Poikilocytosis
Basophil	Blast	Microcytosis
Atypical Lymphocyte	Target Cell	Macrocytosis
Other	Sickle Cell	Platelet Clumps

Reviewed on \_\_\_\_\_ by \_\_\_\_\_ Signature :

**QC - Control Run Report**

Run Date 02/10/2023 08:22:59 PM


Operator ABX

Name HIGH CONTROL  
Level High  
Lot number PX439H


Sample ID PX439H  
Exp. date 03/31/2023

			Range
RBC	5.22	10 <sup>6</sup> /μL	4.99 - 5.49
HGB	15.9	g/dL	15.4 - 16.6
HCT	48.3	%	46.0 - 51.0
MCV	92.6	μm <sup>3</sup>	87.5 - 97.5
MCH	30.4	pg	28.0 - 33.0
MCHC	32.9	g/dL	30.0 - 36.0
RDW-CV	13.0	%	10.5 - 18.5


**Alarms**



			Range
PLT	525	10 <sup>3</sup> /μL	448 - 548
MPV	9.4	μm <sup>3</sup>	7.5 - 11.5



			Range		Range
WBC	18.39	10 <sup>3</sup> /μL	16.10 - 20.50		
	#			%	
NEU	13.21		11.07 -	71.8	60.9 - 80.9
LYM	3.79		2.29 - 5.29	20.6	12.7 - 28.7
MON	0.60		0.00 - 1.32	3.3	0.0 - 7.2
EOS	0.70		0.00 - 1.28	3.8	0.0 - 7.0
BAS	0.09		0.00 - 0.48	0.5	0.0 - 2.6



**Slide Review**

Neutrophil	Myeloblast	Anisocytosis
Lymphocyte	Promyelocyte	Hypochromia
Monocyte	Myelocyte	Polychromasia
Eosinophil	Metamyelocyte	Poikilocytosis
Basophil	Blast	Microcytosis
Atypical Lymphocyte	Target Cell	Macrocytosis
Other	Sickle Cell	Platelet Clumps

Reviewed on \_\_\_\_\_ by \_\_\_\_\_ Signature :



**QC - Control Run Report**

Run Date 02/10/2023 08:24:48 PM

Operator ABX

Name LOW CONTROL

Sample ID PX439L


Level Low

Exp. date 03/31/2023

Lot number PX439L


			Range
RBC	2.36	10 <sup>9</sup> /μL	2.22 - 2.54
HGB	5.9	g/dL	5.6 - 6.4
HCT	18.6	%	16.8 - 19.8
MCV	79.0	μm <sup>3</sup>	72.0 - 82.0
MCH	25.1	pg	23.2 - 27.2
MCHC	31.8	g/dL	29.7 - 35.7
RDW-CV	14.9	%	10.0 - 18.0

**Alarm**  
Control failed  
EOS# above tolerance




RBC

			Range
PLT	77	10 <sup>9</sup> /μL	51 - 91
MPV	8.9	μm <sup>3</sup>	6.9 - 10.9



PLT

			Range		Range
WBC	3.14	10 <sup>9</sup> /μL	2.40 - 3.20		
	#		Range	%	Range
NEU	1.51		0.97 - 1.67	47.9	37.0 - 57.0
LYM	1.14		0.77 - 1.43	36.2	27.4 - 51.4
MON	0.19		0.00 - 0.42	6.2	0.0 - 15.0
EOS	0.29		0.00 - 0.28	9.3	0.0 - 10.0
BAS	0.01		0.00 - 0.06	0.4	0.0 - 2.2



DIF

**Slide Review**

Neutrophil	Myeloblast	Anisocytosis
Lymphocyte	Promyelocyte	Hypochromia
Monocyte	Myelocyte	Polychromasia
Eosinophil	Metamyelocyte	Poikilocytosis
Basophil	Blast	Microcytosis
Atypical Lymphocyte	Target Cell	Macrocytosis
Other	Sickle Cell	Platelet Clumps

Reviewed on \_\_\_\_\_ by \_\_\_\_\_ Signature :