



UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY

CALIBRATION CERTIFICATE

Format No.: UCCS-QF-7.8.1		Certificate No.: UCCSDB20230213-001			
Group: Acceleration & Speed		Discipline: Mechanical			
Customer's Name & Address M/s. Dr. Bharti Path Lab Centre 1, Opp. PNB Bank, Mohna Road, Ballabhgarh		Page No.: 1 of 1	Calibration Performed: At Site		
		Date of Receipt: 13-02-2023			
		Service Request No.: UCCSDBSRN20230213-001			
		Certificate Issue Date: 15-02-2023			
		Date of Calibration: 14-02-2023			
Next Due Date (Suggested by Customer): 13-02-2024					
Equipment Description (UUC) & Identification					
Name of the Instrument: RPM Measurement of Centrifuge Machine		Make: REMI			
Range: Up to 3500 RPM		Model: -----			
Least Count: As per results		Sr. No.: -----			
Condition of Instrument: Ok		I.D. No.: CFM-01			
Master Equipment Used		Location: -----			
Sr. No.	Instrument Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Digital Tachometer	2.5 To 99999 RPM & UCCS/CAL/014	Asian Technology	AT22000000095	06-03-2023
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 25 ± 5 °C, Humidity: 50 ± 10 %RH			UCCS 03 SOP 033	----	

CALIBRATION RESULTS

Sr. No.	Set Knob Position (UUC)	Observed Ref. Standard Value (RPM)	Expanded Uncertainty (±)
1	1	1497	3 % of rdg
2	2	2013	
3	3	2495	
4	4	3011	
5	5	3492	

End of Certificate

Amit
Calibrated By

Amit Kumar

Calibration Engineer

Reviewed & Approved By
M. Mahato
Chief Executive Officer (CEO)

- Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
3. Total liability of our Organization is limited to the invoiced amount.
4. Calibration Certificate without signature are not valid.
5. All our certificates are metrological traceable.
6. The results relate only to the item calibrated.
7. UUC: Unit Under Calibration.
8. Expanded Uncertainty at 95 % confidence level with coverage factor K=2.

Working : W.Z.1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046

Regd. Office : R.Z.56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045

M.: 9958445875, 9313768426, 9999761885 E-Mail : uniquecalibrationservices@gmail.com