



## CALIBRATION REPORT

To  
**Health Horizon Diagnostics  
Pune.**

Date: 04 Mar 2023.

Dear Sir,

As desired by you please find enclosed herewith the report,

Sr. No	56879
Name of the System	INSTALYTE (NA/K/CL)
Calibrated On	01 Mar 2023
Calibration Due On	29 Feb 2024
Calibrated by	Mr. Amurt Kulkarni
Last PM Date	01 Mar 2023

Checked instrument calibration values for-

Na- 58.54

K- 60.00

Cl- 52.00

All values are with in Range

Checked instrument in all respect found ok.

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,  
For **TRANSASIA BIO-MEDICALS LTD.**

  
**Amurt Kulkarni  
ASM**

