

External Quality Assurance Scheme - Print Monthly Summary

7590

Lab No

2022

Year

Dr. UGPHC, Kallur

December

Biochemistry III

Result Entered Date : 07/12/2022

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
				CV	SD	SDI			
GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1844	239.20	6.25	14.96	234 mg/dL	-0.35	0.70	
Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	60	52.81	12.82	6.77	50 mg/dL	-0.42	1.75	
Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.91	8.24	0.32	3.6 mg/dL	-0.96	0.02	
CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1766	114.94	5.41	6.21	118 mg/dL	0.49	0.30	

Interpretation
Excellent.
Good.
Accept with caution. Warning Signal.
Unacceptable performance. Action Signal.

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****