



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



MONTHLY SUMMARY REPORT - DECEMBER 2022

Lab Name Govt. PHC, Nallampalli Lab No 12116
 Constituent Group Chemistry III Date of Result Entered : 17/12/2022
 PT item Lyophilized Serum Date of Report Published : 05/01/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	82	234.68	7.08	16.62	224 mg/dL	-0.64	3.67
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	76	3.97	7.78	0.31	3.4 mg/dL	-1.84	0.07
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1766	114.94	5.41	6.21	134 mg/dL	3.07	0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Nallampalli
 THE MEDICAL OFFICER, NALLAMPALLI PHC, NALLAMPALLI, DHARMAPURI DIST.
 Nallampalli
 TAMILNADU636807

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

J. Nagamani
 MEDICAL OFFICER
 Govt. Primary Health Centre
 NALLAMPALLI
 Dharmapuri (Dist)