

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Nallampalli			Details About Robust Analysis						
Lab No	12116			Detail About Monthly Summary						
Month	November			Detail about SDI						
Year	2022									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data				<a href="#">All Analyser Result</a>		<a href="#">Print</a>				
				Date of Result Entered : 18/11/2022						
				Date of Report Published : 06/12/2022						
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	79	83.80	7.60	6.37	83 mg/dL	-0.13	1.43
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	70	6.97	6.85	0.48	5.8 mg/dL	-2.45	0.11
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1767	112.60	5.74	6.46	109 mg/dL	-0.56	0.31
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*J. Jaganani*  
**MEDICAL OFFICER**  
 Govt. Primary Health Centre  
 NALLAMPALLI  
 Dharmapuri Dt