

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Nallampalli	Details About Robust Analysis
Lab No	12116	Detail About Monthly Summary
Month	October	Detail about SDI
Year	2022	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 22/10/2022

Date of Report Published : 03/11/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	75	343.89	0.89	23.70	318 mg/dL	-1.09	5.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	73	1.38	10.37	0.14	1.2 mg/dL	-1.26	0.03
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	109.51	6.27	6.87	97 mg/dL	-1.82	0.33

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

M. J. JAMANI
MEDICAL OFFICER
Govt. Primary Health Cent.
NALLAMPALLI
Dharmapuri (Dr)