

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Nallampalli		Details About Robust Analysis							
Lab No	12116		Detail About Monthly Summary							
Month	September		Detail about SDI							
Year	2022									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data										
			All Analyser Result	Print						
Date of Result Entered : 20/09/2022										
Date of Report Published : 05/10/2022										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	81	224.82	6.95	15.62	206 mg/dL	-1.20	3.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	73	3.45	9.05	0.31	2.7 mg/dL	-2.40	0.07
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	112.08	5.53	6.20	101 mg/dL	-1.79	0.29
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

M. Janani
 MEDICAL OFFICER
 Govt. Primary Health Centre
 NALLAMPALLI
 Dharmapuri ID#