



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/02029		MEDICAL DEVICES		Page No:1 of 1
Issue Date		20-02-2023		
SRF No & Date		SRF/23/00070-0005 & 20-02-2023		
Receipt Date		20-02-2023		
Calibration Date		20-02-2023		
Calibration Due		19-02-2024		
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, NALLAMPALLI-636807, DHARMAPURI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description : SEMI AUTO ANALYZER		Make & Model : ROBONIK & PREITEST TOUCH		
Range :		Sr. No : ATCD1020621RBK		
Resolution :		Identification No :		
DUC Condition : SATISFACTORY		Location : LAB		
Environmental Conditions & Calibration Procedure Details				
Environmental Details		Temperature: 25.6°C		Relative Humidity
Calibration Procedure No		SBS/CP/MD/20		54% RH
		Calibration done at		ONSITE
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	TSC/22-23/7400-3	10-08-2023

ELECTRICAL SAFETY

RESULTS


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	101	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000μAfor B,BF,CF	213	7.0
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500μAfor B,BF,CF	212	6.7

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

S. Murali
(Calibration Engineer)
S.MURALI



Authorised Signatory

D.V.S.
Quality Manager
(D.VETRI SELVI)

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.