

## CALIBRATION CERTIFICATE CERTIFICATE NO: SBS/CL/23/02031 MEDICAL DEVICES Page No:1 of 1 Issue Date 20-02-2023 SRF No & Date SRF/23/00070-0007 & 20-02-2023 Receipt Date 20-02-2023 Calibration Date 20-02-2023 Calibration Due 19-02-2024 Customer Name & Address GOVERNMENT PRIMARY HEALTH CENTRE. NALLAMPALLI-636807, DHARMAPURI DISTRICT. Details of Device Under Calibration (DUC) Description MICROSCOPE Make & Model OLYMPUS & CH20iBIME Range Sr.No 1111074 Resolution Identification No **DUC Condition** SATISFACTORY Location LAB Environmental Conditions & Calibration Procedure Details **Environmental Details** Temperature: 26.1° C Relative Humidity 50% RH Calibration Procedure No SBS/CP/MD/19

CALID	CALIBRATION RESULTS						
S.No.	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS	DEVIATION	EXPANDED UNCERTAINTY (±)			
LIGHT INTENSITY							
	LUX	LUX	LUX	%			
1	900	885	15	6.9			

Reference Standards Details

Make/ SI No:

Lutron & S.034042

## REMARKS

S.No Description

Digital Light Meter

CALIBRATION DECLIETO

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards

Calibrated By

Authorised Signatory

ONSITE

JRPM-CCTR-OPT-2022-0100

Validity

12-10-2023

(Calibration Engineer)



Quality Manager
(D.VETRI SELVI

Calibration done at

Certificate No

Chief	Exec	άνε

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

