

To,

Date : 26/09/2022

Bhume Diagnostics

Ahmednagar,

Maharashtra

Kind Attn : Mr. Pradeep,

Calibration Certificate

We hear by certify that the Bio - Rad **D - 10** system **Sr. No.DJ8K150602** Installed in your lab Is calibrated today as under & the performance was found satisfactory.

Flow calibration @ 1.5 ml / min (Flow checked by using volumetric variant sample cups with visible observation)

Thermo module temperature calibration @ 35°C (Temp. Checked by using D - 10 internal inbuilt temp.Sensor)

Detector log value @ 0.020 (Voltage verified by using Multimeter)

The next calibration is due on ; 25/09/2023

For,

Anand Chemiceutics

