






CERTIFICATE OF CALIBRATION

| | | | |
|--|---|---|--|
| FT-Q-25 | | Page 1 of 2 | |
| U.I.R No. | : CC214423000001260F | Date of Issue | : 09-03-2023 |
| Certificate No. | : TVCSPL 23/03/431-01 | Recom. Due Date | : 07-03-2024 |
| Date of Calibration | : 08-03-2023 | Customer Details | : M/s. REGIONAL RESEARCH INSTITUTE OF UNANI MEDICINE, NO.1,WEST MATHA CHURCH ROAD,ROYAPURAM, CHENNAI-600013. |
| Customer Details | | SRF No. | : 431 |
| M/s. REGIONAL RESEARCH INSTITUTE OF UNANI MEDICINE, NO.1,WEST MATHA CHURCH ROAD,ROYAPURAM, CHENNAI-600013. | | Calibrated at | : Lab |
| | | Date of Receipt | : 08-03-2023 |
| | | Cond. On Receipt | : Satisfactory |
| Details of Test Instrument: | | | |
| Description | : Micropipette | Model No. | : Biohit |
| Range | : 5-50µl | Serial No. | : 7049845 |
| Least Count | : 0.5µl | Id. No. | : RRIUMC EQ BC 014 |
| Make | : Erba | Accuracy | : As per Manual |
| Operating Range | : 10µl | Location : Instrumentation Room, Dept Of Bio Chemistry | |
| Details of Standard Used | | | |
| Name | : Certificate No. | Valid upto | Traceability |
| Weighing Machine | : TVCSPL 23/03/416-01 | 05-Mar-24 | TVCSPL, Chennai. |
| Weighing Machine | : TVCSPL 23/02/366-03 | 26-Feb-24 | TVCSPL, Chennai. |
| Work Instruction | : WI-M-03 | Environmental Details : Temperature : 25±2°C Relative Humidity : 50±10 % RH | |
| MECHANICAL CALIBRATION | | | |
| (Volume) | | | |
| Calibration Results | | | |
| 1. Lower Volume : | 5 µl | No. of Measurements : | 10 |
| <input type="text" value="5.037"/> | <input type="text" value="5.045"/> | <input type="text" value="5.079"/> | <input type="text" value="5.071"/> |
| <input type="text" value="5.095"/> | <input type="text" value="5.039"/> | <input type="text" value="5.101"/> | <input type="text" value="5.077"/> |
| <input type="text" value="5.065"/> | <input type="text" value="5.059"/> | | |
| Mean Value : | <input type="text" value="5.067"/> µl | | |
| Error Limits(±) | | | |
| Systematic Error : | <input type="text" value="0.067"/> µl | <input type="text" value="0.125"/> µl | |
| Systematic Error : | <input type="text" value="1.34"/> % | <input type="text" value="2.50"/> % | |
| Random Error : | <input type="text" value="0.02"/> µl | <input type="text" value="0.08"/> µl | |
| Random Error : | <input type="text" value="0.45"/> % | <input type="text" value="1.50"/> % | |
| Measurement Uncertainty : | <input type="text" value="± 0.094"/> µl | | |

Calibrated by :

 Ms.K.Vishnu Priyai
 (Calibration Engineer)

Authorised by:

 Mr. Anand Manoj
 (QM & TM)





ULR No. : CC214423000001260F Page 2 of 2
 Certificate No. : TVCSPL 23/03/431-01

MECHANICAL CALIBRATION

(Volume)

Calibration Results

2. Middle Volume : 25 μ l No. of Measurements : 10

| | | | |
|-------|-------|-------|-------|
| 24.91 | 24.74 | 24.88 | 25.00 |
| 25.08 | 24.84 | 25.03 | 24.98 |
| 24.90 | 25.05 | | |

Mean Value : 24.94 μ l

Error Limits(\pm)

| | | | | |
|--------------------|-------|---------|------|---------|
| Systematic Error : | -0.06 | μ l | 0.50 | μ l |
| Systematic Error : | -0.24 | % | 1.00 | % |
| Random Error : | 0.11 | μ l | 0.20 | μ l |
| Random Error : | 0.43 | % | 0.40 | % |

Measurement Uncertainty : \pm 0.26 μ l

3. Nominal Volume : 50 μ l No. of Measurements : 10

| | | | |
|-------|-------|-------|-------|
| 49.87 | 49.84 | 49.98 | 49.86 |
| 49.92 | 50.08 | 50.00 | 49.89 |
| 50.05 | 49.86 | | |

Mean Value : 49.93 μ l

Error Limits(\pm)

| | | | | |
|--------------------|-------|---------|------|---------|
| Systematic Error : | -0.07 | μ l | 0.50 | μ l |
| Systematic Error : | -0.13 | % | 1.00 | % |
| Random Error : | 0.09 | μ l | 0.20 | μ l |
| Random Error : | 0.17 | % | 0.40 | % |

Measurement Uncertainty : \pm 0.26 μ l

Remarks

1. The reported Expanded Uncertainty is calculated at 95.45 % C.L. with coverage factor $k=2$
2. The above Micropipette was within the error limits

* End of Certificate *

Calibrated by :

Vishnu Priyai
 Ms.K.Vishnu Priyai
 (Calibration Engineer)

Authorised by:

Anand Manoharan
 Mr. Anand Manoharan
 (QM & TM)



...redefining the true value





CERTIFICATE OF CALIBRATION

| | | | |
|---|----------------------------|--------------------|----------------------|
| FT-Q-25 | | Page 1 of 2 | |
| ULR No. | : CC214423000001218F | Date of Issue | : 07-03-2023 |
| Certificate No. | : TVCSPL 23/03/416-01 | Recom. Due Date | : 05-03-2024 |
| Date of Calibration | : 06-03-2023 | SRF No. | : 416 |
| Customer Details | | Calibrated at | : LAB |
| M/s. True Value Calibration Services Pvt. Ltd., | | Date of Receipt | : 06-03-2023 |
| No. 92, S. R. B. Nagar Main Road, | | Cond. On Receipt | : Satisfactory |
| Kolathur, | | | |
| Chennai - 600 099. | | | |
| Details of Test Instrument: | | | |
| Description | : Electronic Micro Balance | Model No. | : MSE 3-6P-000-DM |
| Range | : 0 to 1.1g/2.1/3.1g | Serial No | : 36101007 |
| Least Count | : 0.001mg/0.002mg/0.005mg | Identification No. | : TVCSPL/MECH/WB-011 |
| Make | : Sartorius | Class(OIML) | : Class I |
| Working range | : 1mg to 2g | Location | : MASS LAB 1 |
| Verification Interval(e) | : 0.01mg | | |
| Details of Standard Used : | | | |
| Name | Certificate No. | Valid upto | Traceability |
| E1 Class Standard Weights | TVCSPL 22/05/694-03 | 03-May-23 | TVCSPL, Chennai. |
| Work Instruction | : WI-M-01 | | |
| Environmental Details | Temperature: 25±5 °C | Relative Humidity: | 40-70 % RH |

Calibrated by :
K. Kameswaran
 Mr.K.Kameswaran
 (Calibration Engineer)

Authorised by:
Anand Manoharan
 Mr.Anand Manoharan
 (QM & TM)



...redefining the true value

92, S.R.B Nagar Main Road,
 Chennai - 600 099.
 Tamil Nadu, India.

Ph: 044 - 4281 9208 / Cell: 94440 38069 /
 97102 22422 / 97102 22522 / 97102 22622
 Email : calibrationservices@live.com
 www.truevaluecalibration.com
 CIN No :U29268TN2015PTC103428



Quality is Assured

ULR No. : CC214423000001218F
 Certificate No. : TVCSPL 23/03/416-01

Page 2 of 2

MECHANICAL CALIBRATION
 (Weighing Scale & Balance)
Calibration Results

1. Weighing Error Test

| Sr. No. | Applied Mass (g) | Test Reading (g) | Error (g) | MPE(±)g |
|---------|------------------|------------------|-----------|----------|
| 1 | 0.001002 | 0.001000 | -0.000002 | 0.000005 |
| 2 | 0.002002 | 0.002001 | -0.000001 | |
| 3 | 0.005002 | 0.005000 | -0.000002 | |
| 4 | 0.010001 | 0.010001 | 0.000000 | |
| 5 | 0.020001 | 0.020001 | 0.000000 | |
| 6 | 0.050003 | 0.050001 | -0.000002 | |
| 7 | 0.100003 | 0.100001 | -0.000002 | |
| 8 | 0.200003 | 0.200000 | -0.000003 | |
| 9 | 0.499999 | 0.500000 | 0.000001 | |
| 10 | 1.000007 | 1.000000 | -0.000007 | 0.000010 |
| 11 | 2.000008 | 2.000000 | -0.000008 | |
| 12 | 3.000015 | 3.000000 | -0.000015 | |

2. Repeatability

| Sr. No. | @ Zero (g) | 50% of Range (g) | 100% of Range (g) |
|---------|------------|------------------|-------------------|
| 1 | 0.000000 | 1.000000 | 2.000002 |
| 2 | 0.000000 | 1.000000 | 2.000000 |
| 3 | 0.000000 | 1.000000 | 2.000000 |
| 4 | 0.000000 | 1.000000 | 2.000002 |
| 5 | 0.000000 | 1.000000 | 2.000000 |
| 6 | 0.000000 | 1.000001 | 2.000002 |
| 7 | 0.000000 | 1.000001 | 2.000000 |
| 8 | 0.000000 | 1.000000 | 2.000002 |
| 9 | 0.000000 | 1.000001 | 2.000002 |
| 10 | 0.000000 | 1.000001 | 2.000002 |

Repeatability (SD) ± 0.000001 g

Remarks

- UUC is defined as the Unit Under Calibration
- Expanded uncertainty** ± 0.0000032 g
- The reported uncertainty is at coverage k=2 which correspond to a coverage probability of approximately 95.45% for a normal distribution. The Contribution of uncertainty, originating from the standards(s) & balance(s) used, the weighing process are into account
- M/s.TVCSPL, Chennai General Laboratory Practice are derived from ISO/IEC:17025. It therefore meets the relevant requirements of ISO:9001 when acting as the supplier providing test/calibration result
- The result provided in this certificate is confidential. The calibration certificate shall not be reproduced in part or in full without written approval of M/s.TVCSPL.
- With respect to the above specified requirements the balance was within the limits
- This calibration certificate will not be legal for the purpose of the standard of "Weight & Measure (enforcement) act 2011

* End of Certificate *

Calibrated by :

K. Kameswaran
 Mr.K.Kameswaran
 (Calibration Engineer)

Authorised by:

Anand Manoharan
 Mr.Anand Manoharan
 (QM & TM)



...redefining the true value



CC-2144

CERTIFICATE OF CALIBRATION

| | | | |
|---|---------------------------------|--------------------|---------------------|
| FT-Q-25 | | Page 1 of 2 | |
| ULR No. | : CC214423000001113F | Date of Issue | : 28-02-2023 |
| Certificate No. | : TVCSPL 23/02/366-03 | Recom. Due Date | : 26-02-2024 |
| Date of Calibration | : 27-02-2023 | SRF No. | : 366 |
| Customer Details | | Calibrated at | : LAB |
| M/s. True Value Calibration Services Pvt. Ltd., | | Date of Receipt | : 27-02-2023 |
| No. 92, S. R. B. Nagar Main Road, | | Cond. On Receipt | : Satisfactory |
| Kolathur, | | | |
| Chennai - 600 099. | | | |
| Details of Test Instrument: | | | |
| Description | : Electronic Semi Micro Balance | Model No. | : MSA 225 S-000-DA |
| Range | : 0 to 220g | Serial No | : 28601485 |
| Least Count | : 0.01mg | Identification No. | : TVCSPL/MECH/WB-07 |
| Make | : Sartorius | Class(OIML) | : Class I |
| Working range | : 1mg to 200g | Location | : MASS LAB 1 |
| Verification Interval(e) | : 0.1mg | | |
| Details of Standard Used | | | |
| Name | Certificate No. | Valid upto | Traceability |
| E1 Class Standard Weights | TVCSPL 22/05/694-03 | 03-May-23 | TVCSPL, Chennai. |
| Work Instruction | : WI-M-01 | | |
| Environmental Details | : Temperature: 25±5 °C | Relative Humid | 40-70 % RH |

MECHANICAL CALIBRATION


(Weighing Scale & Balance)

Calibration Results

1. Eccentricity

| Sr. No. | Position of Weights used | Test Reading (g) | Error between centre(1) and Other points (g) |
|---------|--------------------------|------------------|--|
| 1 | 3 | 50.00000 | 0.00001 |
| 2 | 4 | 50.00001 | 0.00000 |
| 3 | 1 | 50.00001 | 0.00000 |
| 4 | 2 | 50.00000 | 0.00001 |
| 5 | 5 | 50.00000 | 0.00001 |

Maximum Error between centre (1) and other points
0.00001 g



Calibrated by :

K. Kameswaran
Mr.K.Kameswaran
(Calibration Engineer)

Authorised by:

Anand Manoharan
Mr.Anand Manoharan
(QM & TM)





ULR No. : CC214423000001113F
 Certificate No. : TVCSPL 23/02/366-03

MECHANICAL CALIBRATION
 (Weighing Scale & Balance)
Calibration Results

2. Weighing Error Test

| Sr. No. | Applied Mass (g) | Test Reading (g) | Error (g) | MPE(±)g |
|---------|------------------|------------------|-----------|---------|
| 1 | 0.010001 | 0.01000 | -0.000001 | 0.00005 |
| 2 | 0.499999 | 0.50000 | 0.000001 | |
| 3 | 1.000007 | 1.00000 | -0.000007 | |
| 4 | 2.000008 | 2.00000 | -0.000008 | |
| 5 | 5.00001 | 5.00000 | -0.00001 | |
| 6 | 10.00001 | 10.00000 | -0.00001 | 0.00010 |
| 7 | 20.00001 | 20.00000 | -0.00001 | |
| 8 | 50.00002 | 50.00000 | -0.00002 | 0.00015 |
| 9 | 100.00003 | 100.00000 | -0.00003 | |
| 10 | 200.00004 | 200.00000 | -0.00004 | |

3. Repeatability

| Sr. No. | @ Zero (g) | 50% of Range (g) | 100% of Range (g) |
|---------|------------|------------------|-------------------|
| 1 | 0.00000 | 100.00000 | 200.00000 |
| 2 | 0.00000 | 100.00000 | 200.00000 |
| 3 | 0.00000 | 100.00000 | 200.00001 |
| 4 | 0.00000 | 100.00000 | 200.00001 |
| 5 | 0.00000 | 100.00000 | 200.00001 |
| 6 | 0.00000 | 100.00000 | 200.00001 |
| 7 | 0.00000 | 100.00001 | 200.00001 |
| 8 | 0.00000 | 100.00001 | 200.00001 |
| 9 | 0.00000 | 100.00001 | 200.00000 |
| 10 | 0.00000 | 100.00000 | 200.00000 |

Repeatability(SD)

± 0.000005 g

Remarks

- UUC is defined as the Unit Under Calibration
- Expanded uncertainty** ± 0.000025 g
- The reported uncertainty is at coverage k=2 which correspond to a coverage probability of approximately 95.45% for a normal distribution. The Contribution of uncertainty, originating from the standards(s) & balance(s) used, the weighing process are into account
- M/s.TVCSPL, Chennai General Laboratory Practice are derived from ISO/IEC:17025. It therefore meets the relevant requirements of ISO:9001 when acting as the supplier providing test/calibration result
- The result provided in this certificate is confidential. The calibration certificate shall not be reproduced in part or in full without written approval of M/s.TVCSPL
- With respect to the above specified requirements the balance was within the limits
- This calibration certificate will not be legal for the purpose of the standard of "Weight & Measure (enforcement) act 2011

* End of Certificate *

Calibrated by :
K. Kameshwaran
 Mr.K.Kameswaran
 (Calibration Engineer)

...redefining the true value

Authorised by:
Anand Manoharan
 Mr.Anand Manoharan
 (QM & TM)

