

**OFFICE OF THE MEDICAL SUPERINTENDENT
DEPARTMENT OF MICROBIOLOGY
CIVIL HOSPITAL: AIZAWL, MIZORAM**

Letter No. D.28016/03/2023/Micro/023

Date: 14 .03.2023

To,

**The In-charge
ICTC, Christian Hospital Serkawn
Lunglei, Mizoram**

Subject: HIV Proficiency Test (Round – I, EQA : 2022-2023) Feedback

Sir/Madam,

*Please find enclosed herewith the Report of EQAS Proficiency Testing Report , **Round- I, EQA: 2022 - 2023** distributed on **28th of February, 2023**. Your Centre has detected the Panel Sera correctly and has **Scored 100% result**.*

Looking forward for your future Participation in the HIV EQAS Programme.

Thanks and regards.

Dated: Aizawl the 14th March, 2023

Your's faithfully,

per Loh

**(Dr. LALHMINGMAWII)
In-charge
State Reference Laboratory
cum ICTC
Microbiology Department
Civil Hospital, Aizawl: Mizoram**

Date: 14/03/2023

PROFICIENCY TEST REPORT (SRL to ICTC)

Round-I/ II : I- Round EQA : 2022-2023

Name of Proficiency Testing Provider:
NACO-NRL, ICMR-NICED

Date of Panel Distribution:
28th Feb. 2023

Name of Participating Laboratory:
ICTC, Christian Hospital Serkawn

Date of Result Received:
07th March 2023

EQAS Laboratory I.D: MZ - 16

| Sample Code | True Status | ICTC/Blood Bank | Matching Yes/No | Score |
|-------------|------------------|------------------|-----------------|---------------|
| 22/I - A16 | Reactive(HIV-I) | Reactive(HIV-I) | Yes | 100% Matching |
| 22/I - B16 | Non- Reactive | Non- Reactive | Yes | |
| 22/I - C16 | Non- Reactive | Non- Reactive | Yes | |
| 22/I - D16 | Non- Reactive | Non- Reactive | Yes | |

Keyword: R: Reactive NR: Non-Reactive

Remarks: Satisfactory.

Please note that the boxes given on the final interpretation column were leaved blank for all the Negative samples. Fill in all the revelant spaces provided legibly on the Reporting Format.

Name of the Laboratory Incharge:

Dr. LALHMINGMAWII

ha Lahr
(Signature)