



arkray healthcare Pvt. Ltd
Plot no. 336, 338, 340, Road no. 3, GLDC,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL: <http://www.arkray.co.in>

Certificate Of Calibration

This is to certify that our service engineer Mr. Sachin Singh has attended **U120 Urine analyzer** (Sr.no: 197K10007EA) on date 28/09/2022 installed at **REDCLIFFE LIFETECH PVT. LTD. Patna - 800020**

During his visit, he has **done Calibration & checked performance** of the instrument as listed below

U120 Urine analyzer Serial no. 197K10007EA
Self-check O.K

Main control unit	Pass
Optical sensor	Pass
Test accuracy	Pass
Mechanism	Pass
Excess Light	Pass

On basis of above mention checks, the instrument is functioning satisfactory.
This is valid up to 27/09/2023. If any major breakdown occur instrument to be re-validated.

Best Regards,

Sachin Singh (SE)
Arkray Healthcare Pvt Ltd

Date:17-03-2023 11:44

Operator: 04

Level 2 Pass

No.200126

LEU	3+	500 Leu/uL
NIT	+	pos
URO	3+	140 umol/L
PRO	3+	3.0 g/L
pH		6.5
BLO	3+	200 Ery/uL
SG		1.020
KET	2+	4.0 mmol/L
BIL	3+	70 umol/L
GLU	3+	60 mmol/L

QC Passed

Date:17-03-2023 11:40

Operator: 04

Level 1 Pass

No.200124

LEU	-	neg
NIT	-	neg
URO	-	3.5 umol/L
PRO	-	neg
pH		5.5
BLO	-	neg
SG		1.025
KET	-	neg
BIL	-	neg
GLU	-	neg

Reforn
17-03-2023

QC
17-3-2023