



Doc No: CC/EL120/12.03.2022

AGD Biomedicals Pvt.Ltd
AGD Biomedicals (P) Ltd. Mehta
Trade Center, No.1, Sir Mathuradas
VasANJI Rd, Shivaji Colony, Andheri
East, Mumbai, Maharashtra 400059

Calibration Certificate

Customer Name : Antlia Lab and Diagnostics Pvt. LTD Ranchi

Equipment Name : Clinical Chemistry Analyser

Model Name : AGD2020 Clinical Chemistry Analyser

Serial Number : AB2109037

Calibration Date : 11.03.2023

Validity of Calibration : 10.03.2024

Authorized Signatory : Rakesh Kumar

AGD Biomedicals (P) Ltd

Note : Recommended to perform calibration every 6 months.

Calibration Passed

Model Name : AGD 2020 Clinical Chemistry Analyser

Date of Performance : 11.03.2023

Performance :

Calibration Performance										
Parameters	S1	S2	S3	Mean	SD	CV	Target	Range	1 SD	2 SD
Glucose	100	99	100	99.67	0.58	0.58	100.00	98-102	1.00	1.00
Tryglycerides	150	149	150	149.67	0.58	0.39	150.00	148-152	1.00	2.00
SGOT	50	51	52	51.00	1.00	1.96	50.00	48-52	1.00	2.00
Creatinine	2.2	1.98	2.01	2.06	0.12	5.78	2.20	2.15-2.25	0.25	0.25

Rohit Kumar
Name and Signature of Customer
Seal & Date
Rohit Kumar



MR. RAKESH KUMAR
Name and Signature of Engineer

Seal & Date

