



Doc No: CC/EL120/12.03.2022

AGD Biomedicals Pvt. Ltd
AGD Biomedicals (P) Ltd. Mehta
Trade Center, No.1, Sir Mathuradas
Vasnaji Rd, Shivaji Colony, Andheri
East, Mumbai, Maharashtra 400059

Calibration Certificate

Customer Name : Antlia Lab and Diagnostics Pvt. LTD Ranchi

Model Name : EL-120 Electrolyte Analyzer

Serial Number : ED211001

Calibration Date : 11.03.2023

Validity of Calibration : 10.03.2024

Authorized Signatory Rakesh Kumar

AGD Biomedicals (P) Ltd

Note : Recommended to perform calibration every 6 months.

Calibration Passed

Model Name : EL 120
Date of Performance : 11.03.2022

Calibration Performance										
Parameters	S1	S2	S3	Mean	SD	CV	Target	Range	1 SD	2 SD
Sodium	141	142	142	141.67	0.58	0.41	142.00	140-144	3.00	6.00
Potassium	4.7	4.7	4.7	4.70	0.00	0.00	4.68	4.65-4.71	0.20	0.40
Chloride	101	99	99	99.67	1.15	1.16	100.00	98-102	0.15	0.30

Rohit Kumar
Name and Signature of Customer
Seal & Date
Rohit K.



Mr. Rakesh Kumar
Name and Signature of Engineer
Seal & Date

