



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 5341

Distribution No.: 158-N

Month/Year: February/2023

Instrument ID: Yumizen H 550 - 201YAXH03629

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 28-02-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	1.2	1.02	2.22	14.94	0.0360	-14.21	0.18	0.1	0.0080	0.77
RBC x10 <sup>6</sup> /µl	1	4.01	3.99	8	7.31	0.0090	2.91	0.02	0.04	0.0020	-0.54
Hb g/dl	1	12.8	12.6	25.4	24.9	0.0270	0.82	0.2	0.1	0.0080	0.79
HCT%	1	38.2	38.2	76.4	78.15	0.1710	-0.33	0	0.4	0.0250	-0.90
MCV-fl	1	95.7	95.2	190.9	214.85	0.4180	-1.80	0.5	0.3	0.0230	0.54
MCH-Pg	1	32	31.4	63.4	68.1	0.0800	-2.24	0.6	0.3	0.0220	0.81
MCHC-g/dl	1	33.4	33	66.4	63.4	0.1470	0.67	0.4	0.3	0.0200	0.27
Plt. x10 <sup>3</sup> /µl	1	179	169	348	319	1.68	0.56	10	5	0.32	0.96
Retic %	2	5	4	9	10	0.18	-0.20	1	0.5	0.03	0.84

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=4 , Poly=6 L=89, E=0, Mono/Promono=3 , B1=2 P.M.=0, Mye=0, Meta=0, Other=Atypical Lymphoid cells seen	Lympho: 49-73, Blast: 5-35, Poly:4-10, nRBC/Eos/Baso/Mono/Myelo/Meta/ Promyelo: 0-5		
RBC Morphology	3	Normocytic Normochromic, Spherocytes	Predominantly: Normocytic/ Normochromic, Moderate: Anisocytosis, Microcytic		
Diagnosis	3	Aleukemic Leukemia	Acute Leukemia (AL)		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 158--N	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	342	340	80	85.59	4.71	3.53	15.29	10.88
<b>RBC x10<sup>6</sup>/µl</b>	1	342	342	88.3	90.35	6.73	3.51	4.97	6.14
<b>Hb g/dl</b>	1	342	342	86.55	85.96	6.73	6.73	6.72	7.31
<b>HCT%</b>	1	342	340	95.88	87.94	3.53	5.29	0.59	6.77
<b>MCV-fl</b>	1	342	340	97.35	92.35	1.76	2.35	0.89	5.3
<b>MCH-Pg</b>	1	342	340	86.47	88.24	8.82	6.47	4.71	5.29
<b>MCHC-g/dl</b>	1	342	340	95.59	90.88	3.24	4.71	1.17	4.41
<b>Plt. x10<sup>3</sup>/µl</b>	1	342	340	95	90	3.24	3.82	1.76	6.18
<b>ReticCount%</b>	2	342	265	94.34	89.81	4.53	5.66	1.13	4.53
<b>PS Assessment</b>	3	342	244	Satisfactory :81.59%, Borderline Sat. :11.40%, Unsatisfactory :7.01%					

**\*Comments:**

1). **Among Lab (EQA) : CBC result for WBC unacceptable, may be due to random/human error.PS Diagnosis partially correct, remaining results acceptable**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----