

Name	Jagannath Choudhary
Department	Healthcare – Service
Telephone	+91-022 3370 0725
Fax	+91-022 3370 0654
Mobile	+91- 09769108282
E-mail	Jagannath.choudhary@siemens-healthineers.com

Calibration Certificate.

This is to certify that installed at Kamal Diagnostics LLP, Pune. APK 300 Instrument ID 6640000049368 Serial number 908020059 has been duly calibrated on March 02, 2023 and the same was verified with the following tests.

System Tests results:

Details	Remark
● Temperature	Passed
● Stray light	Passed
● Noise	Passed
● Stability	Passed
● Tip pump	Passed
● Level detection	Passed
● Washer hydraulics	Passed
● Washer	Passed
● Clot detector	Passed

Verified the Calibrations and alignment visually. Performed Tests , all are passed. Calibration of the instrument was carried out as per the recommendation. Next Calibration will be performed on March 01, 2024.

Installation Date: 12/ 03/ 2022	Software Revision #: 2.6.1s
Install Technician Name: Ajinkya kubade	Signature:

Siemens Healthcare Private Limited



Jagannath Choudhary
Regional Customer Care Manager – West & West Central

Siemens Healthcare Private Limited

Unit No.9A, 9th Floor, North Tower,
 Godrej One, Pirojshanagar,
 Vikhroli (E), Mumbai - 400 079

Tel.: +91 022-3370 0600
 Fax: +91 022-3370 0604
 Email: hc_contact.india@siemens-ealthineers.com
 www.siemens.co.in

Registered office: Unit No. 9A, 9th Floor, North Tower, Godrej One, Pirojshanagar, Eastern Express Highway, Vikhroli (E), Mumbai - 400 079; Corporate Identity number: U74999MH2015PTC264859
 Sales Offices: Ahmedabad, Bengaluru, Bhopal, Chandigarh, Chennai, Coimbatore, Gurgaon, Hyderabad, Jaipur, Kochi, Kolkata, Lucknow, Mumbai, Nagpur, Patna, Pune.