

CALIBRATION REPORT

Centrifuge Table top



Report No:	CAL-AR-2023-02-0970	Date:16-03-2023	Page 01 of 02
------------	---------------------	-----------------	---------------

Customer Details			
Health Facility	Tomo Riba State Hospital		
State	Arunachal Pradesh	District	Papum Pare

Medical Equipment Details			
ME Number	ARZ1PAPDH0080383	Manufacturer	Remi
ME Name	Centrifuges, Tabletop (Centrifuge)	Model	C85416
Department	SrI-Old Building Road Side	Serial No	00LC-2756
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	14-03-2023	Due for Calibration	13-03-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	57%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0191	DIGITAL TACHOMETER	AMBROBE	TACH-10	19020097	23-11-2023

Performance Results						
No	Parameters	Unit	Set Value	MeasuredValue	Acceptable Range	Remarks
1	Speed	rpm	1500	1525	1425 to 1575	Pass
2	Speed	rpm	2500	2540	2375 to 2625	Pass
3	Speed	rpm	3500	3600	3325 to 3675	Pass

CALIBRATION REPORT

Centrifuge Table top



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval.

Calibration Status

Passed

Failed

Calibrated By

C. Vinath Raja

C. Vinath Raja

Authorised Signatory

Rahul Kumar Kasyap

Rahul Kumar Kasyap