CALIBRATION REPORT





Report No:	CAL-AR-2023-02-0966	Date: 16-03-2023	Page 01 of 02

Customer Details			
Health Facility	Tomo Riba State Hospital		
State	Arunachal Pradesh	District	Papum Pare

Medical Equipment Details						
ME Number	ARZ1PAPDH0080375	Manufacturer	Godrej			
ME Name	Refrigerators, Laboratory (Lab refrigerator)	Model	NA			
Department	lctc-Tb Dept.Gate No.1	Serial No	10719			
Calibrated at	On Site	Condition of ME	Good			
Date of Calibration	14-03-2023	Due for Calibration	13-03-2024			

Environmental Conditions					
Temperature in °C	20°C	Humidity in % Rh	58%		

Deta	Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity	
1	TRI-TE-0018	DIGITAL THERMOMETER	METRAVI	DTM- 900	15061597 5	25-08-2023	

Peri	Performance Results						
No	Parameter	Set Temp	Unit	Measured Value	Acceptable Range	Remarks	
1	Temperature	2°C	°C	2.2	1.5 to 2.5°C	Pass	
2	Temperature	4°C	°C	4.2	3.5 to 4.5°C	Pass	
3	Temperature	6°C	° C	6.2	5.5 to 6.5°C	Pass	

CALIBRATION REPORT





Conditions of Issue of report 1. The Standards used are traceable to National standards. 2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited. Calibration Status Passed ☑ Failed □ Calibrated By Authorised Signatory Rahul Kumar Kasyap