



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Name Govt. PHC, Nagavedu **Lab No** 13018
Instituent Group Chemistry III **Date of Result Entered :** 14/02/2023
Item Lyophilized human serum based **Date of Report Published :** 08/03/2023

No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	154 mg/dL	-0.11	0.47
	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	30.39	12.93	3.93	32 mg/dL	0.41	0.59
	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3.17 mg/dL	-0.75	0.01
	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	114 mg/dL	0.20	0.28

SDI Range	Interpretation
SDI in -1.00 to +1.00	Excellent.
SDI in ±1.01 to ±2.00	Good.
SDI in ±2.01 to ±2.99	Accept with caution. Warning Signal.
SDI beyond ±3.0	Unacceptable performance. Action Signal.

ADDRESS :
 Govt. PHC, Nagavedu
 Nagavedu
 Vellore
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****