



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name: Govt. PHC, Nagavedu

Constituent Group: Chemistry III

PT Item: Lyophilized Serum

Lab No: 13018

Date of Result Entered: 12/12/2022

Date of Report Published: 05/01/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		CV	SD	CV	SD
						CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1844	239.20	6.25	14.96	218	mg/dL	-1.42	0.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.91	8.24	0.32	4.2	mg/dL	0.90	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1766	114.94	5.41	6.21	126	mg/dL	1.78	0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Nagavedu

Nagavedu

Vellore

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not subcontract any components

***** End of Report *****

@.vetowly

மருத்துவ அலுவலர்,
அரசு ஆரம்ப சுகாதார நிலையம்,
நாகவேடு-631 002.
இராணிப்பேட்டை மாவட்டம்.