



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2022



Name: Govt. PHC, Nagavedu Lab No: 13018
 Constituent: Chemistry III Date of Result Entered: 14/11/2022
 Sample: Lyophilized Serum Date of Report Published: 06/12/2022

Sample	Method / Principle	Analyzer / Instrument	No. of Determinations	Mean	SD	CV	Target	Delta	Delta 2
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	82.95	8.62	7.15	75 mg/dL	1.11	0.33
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1843	6.98	7.13	0.50	7.6 mg/dL	1.25	0.02
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1767	112.60	5.74	6.46	108 mg/dL	-0.71	0.31

SDI Range	Interpretation
n -1.00 to +1.00	Excellent.
n ±1.01 to ±2.00	Good.
n ±2.01 to ±2.99	Accept with caution. Warning Signal.
and ±3.0	Unacceptable performance. Action Signal.

ADDRESS :
 PHC, Nagavedu
 Vellore
 TNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

Elaja

மருத்துவ அலுவலர்,
 அரசு ஆரம்ப சுகாதார நிலையம்,
 நாகவேடு-631 002,
 இராணிப்பேட்டை மாவட்டம்.