



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - OCTOBER 2022



PC-1024

Lab Name: Govt. PHC, Nagavedu Lab No: 13018  
Constituent Group: Chemistry III Date of Result Entered: 13/10/2022  
Item: Lyophilized Serum Date of Report Published: 03/11/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1837	346.52	7.17	24.84	301 mg/dL	-1.83	1.16
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1831	1.36	10.34	0.14	1.4 mg/dL	0.28	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	109.51	6.27	6.87	112 mg/dL	0.36	0.33

SDI Range	Interpretation
within -1.00 to +1.00	Excellent.
within ±1.01 to ±2.00	Good.
within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Nagavedu  
Nagavedu  
AMILNADU

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Relax*  
மருத்துவ அலுவலர்,  
அரசு ஆரம்ப சுகாதார நிலையம்  
நாகவேடு-631 002.  
இராணிப்பேட்டை மாவட்டம்.