



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - **SEPTEMBER 2022**



Name **Govt. PHC, Nagavedu**

Lab No 13018

Constituent Chemistry III

Date of Result Entered : 17/09/2022

Item Lyophilized Serum

Date of Report Published : 05/10/2022

Test Name	Method	Instrument	Lab No	Result	Target	Delta	Delta 2	Delta 3	Delta 4
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1846	223.49	6.67	14.90	214 mg/dL	-0.64	0.69
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1834	3.49	9.05	0.32	3.4 mg/dL	-0.28	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	112.08	5.53	6.20	106 mg/dL	-0.66	0.29

SDI Range	Interpretation
SDI in -1.00 to +1.00	Excellent.
SDI in ±1.01 to ±2.00	Good.
SDI in ±2.01 to ±2.99	Accept with caution. Warning Signal.
SDI in ±3.0 and above	Unacceptable performance. Action Signal.

ADDRESS :
Govt. PHC, Nagavedu
Nagavedu
Vellore
TAMILNADU

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Subja

மருத்துவ அலுவலர்,
அரசு ஆரம்ப சுகாதார நிலையம்,
நாகவேடு-631 002.
இராணிப்பேட்டை மாவட்டம்.