



# Om Sales Corporation

## CALIBRATION CERTIFICATE FOR NULYTE

This is to certify that, Nulyte Na/K/Cl is a fully automated Electrolyte Analyzer and it takes one-point calibration after every sample and two-point calibration after every Eight hours automatically. Therefore, further manual calibration is not required for this equipment.

*Nulyte Electrolyte* Analyzer, **serial no-OEM-377** installed at Bhume Pathology Laboratory Pvt.Ltd., Ahmednagar is serviced & calibrated on **05/11/2022**.

This instrument is calibrated through Calibration solution present in Solution Pack.

This is to certify that the above instrument is working satisfactory & Calibration is done for above Parameters.

The Next Scheduled Calibration is due on **04/11/2023**.

For Om Sales Corporation

Yash Ghanekar

(Name and Designation of the Engineer)

Mobile No.: 9307475255

*Aman Bunglow, Ground floor S. No 5A/1B, CTS no 696, Plot no 41, Lane no 2, Vidisha Society, Hingane*

*Karve Nagar, Pune 411052, Contact no-7798050099 email:admin@omsales.org*