



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Lab Name	DIALIFE DIABETES THYROID AND ENDOCRINE CENTRE	Lab No	4156
Constituent Group	HbA1c	Date of Result Entered :	21/02/2023
PT Item	Lyophilized human whole blood based	Date of Report Published :	08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	139	6.76	8.40	0.57	7 %	0.42	0.10

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Comments: Lab value shows excellent correlation with peer group mean

LAB ADDRESS :

DIALIFE DIABETES THYROID AND ENDOCRINE CENTRE
 No.962,BEL LAYOUT , VIDYARANYAPURA MAIN ROAD , VIDYARANYAPURA
 BENGALURU
 KARNATAKA560097

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

DIALIFE
 Adding Sweetness to Your Life
 DIABETES, THYROID & ENDOCRINE CENTRE
 # 962, 1st Floor, Above Cinarathin
 BEL Layout, Vidyananyapura Main Road
 Adyaranyapura, Bangalore

Prasad