

VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. Block PHC, Madanur	<a href="#">Details About Robust Analysis</a>
<b>Lab No</b>	3075	<a href="#">Detail About Monthly Summary</a>
<b>Month</b>	February	<a href="#">Detail about SDI</a>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)   [Print](#)

Date of Result Entered : 22/02/2023

Date of Report Published : 08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1868	155.11	5.56	10.18	159 mg/dL	0.38	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1872	3.40	9.00	0.31	3.9 mg/dL	1.63	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1788	112.79	5.23	5.90	127 mg/dL	2.41	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out