



Abbott Healthcare Pvt. Ltd.
Floor 13-19, Godrej BKC, Plot C-68
G Block, Bandra Kurla Complex
Near MCA Club, Bandra East
Mumbai - 400051. India

Registered Office:
4, Corporate Park, Sion
Trombay Road, Mumbai
- 400071. India

T : +91 22 3816 2000 / 1010
F : +91 22 3816 2400 / 2401
Email : webmaster@abbott.co.in
Site : www.abbottdiagnostics.com
CIN : U24200MH1997PTC104834

Preventive Maintenance & Calibration Certificate

This is to inform that the Abbott Architect ci4100 fully automated Biochemistry & immunology Analyzer bearing the serial number S/N:C402951&i1sr62507 Installed at Telangana Diagnostics, Kothagudem has been performed with the PM & Hardware Calibration of the instrument. The result are within the acceptable limit. The Hardware Calibration & PM was performed on 23rd May 2022.

The next Preventive Maintenance and Hardware Calibration due is on 22nd May 2023

For Abbott Healthcare Pvt Ltd.

A small, square image showing a handwritten signature in blue ink, which appears to be 'Mahesh Lokam'.

Mahesh.Lokam

(Technical support specialist)

Procedure Report

Procedure: 1114 Carrier Transport Calibration

Procedure version: 13

Date: 23.05.2022

Module no.: 0

Time: 13:34:44

Serial no.: RSH02951

Status: Completed

Operator ID: FSE

Results:

Position	Result
X at bay 1	2884
X at bay 2	7892
X at bay 3	12899
X at bay 4	17907
X at bay 5	27739
X at bay 6	32745
X at bay 7	37751
X at bay 8	42759
Theta at aspiration	168
Z at positioner pocket 1	8052
X at positioner pocket 1	14105
Z at positioner pocket 2	8059
X at positioner pocket 2	9077
Z at bay 1	7993

Procedure version: 13

Module no.: 0

Serial no.: RSH02951

Operator ID: FSE

Date: 23.05.2022

Time: 13:34:44

Status: Completed

Results:

Position	Result
Z at bay 2	7960
Z at bay 3	7953
Z at bay 4	7972
Z at bay 5	8105
Z at bay 6	8117
Z at bay 7	8064
Z at bay 8	8025
Latch actuator position	44145
X at bottle rotator	28323
Z at bottle rotator	2337

Procedure Report

Procedure: 1120 Sample Pipettor Calibration

Procedure version: 12

Date: 23.05.2022

Module no.: 1

Time: 13:58:39

Serial no.: c402951

Status: Completed

Operator ID: FSE

Results:

Calibration Point	Theta	Z	Carousel	Carrier Position
-------------------	-------	---	----------	------------------

Reaction Carousel	0	30	1	NA
Wash Cup	2	NA	NA	NA
Sample Handler	1	32	NA	62

Calibration passed.

Procedure Report

Procedure: 1122 R2 Pipettor Calibration

Procedure version: 9

Date: 23.05.2022

Module no.: 1

Time: 14:22:18

Serial no.: c402951

Status: Completed

Operator ID: FSE

Results:

Calibration Point	Theta	Z	Carousel
Wash Cup	1	NA	NA
Reaction Carousel	0	NA	1
Outer Segment	2	0	22
Inner Segment	2	1	50

Calibration passed

Procedure Report

Procedure: 1121 R1 Pipettor Calibration

Procedure version: 10

Date: 23.05.2022

Module no.: 1

Time: 14:33:45

Serial no.: c402951

Status: Completed

Operator ID: FSE

Results:

Calibration Point	Theta	Z	Carousel
-------------------	-------	---	----------

Calibration Point	Theta	Z	Carousel
Wash Cup	3	NA	NA
Reaction Carousel	2	NA	1
Outer Segment	0	-3	18
Inner Segment	0	2	15

Calibration passed

Procedure Report

Procedure: 6052 Wash Cuvettes

Procedure version: 4

Module no.: 1

Serial no.: c402951

Operator ID: FSE

Date: 23.05.2022

Time: 13:11:39

Status: Completed

Results:

6052 Wash Cuvettes procedure completed.

Procedure Report

Procedure: 3525 Temperature Status

Procedure version: 4

Date: 23.05.2022

Module no.: 1

Time: 14:39:42

Serial no.: c402951

Status: Completed

Operator ID: FSE

Results:

Processing Area	Temperature	Status
Water Bath	37.05	OK
Reagent Supply Center.	5.20	OK
Instrument Internal.	29.60	OK