




Calibrator Calibration

Instrument: 5 Part Haematology Analyzer		Model: BC -6200		Serial No.: TW-02000920		Date of Cal. 29.07.2022
Mode: OV-WB		Calibrator Lot No. PLUS0722			Exp.Date:05.08.2022	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.57	4.66	14.4	92.3	246	8.58
2.	8.67	4.67	14.3	92.3	245	8.59
3.	8.65	4.64	14.3	91.8	252	8.57
4.	8.85	4.71	14.3	91.7	249	8.43
5.	8.97	4.75	14.3	91.6	247	8.27
6.	8.80	4.72	14.4	91.7	250	8.74
Mean	8.75	4.69	14.3	91.9	248	8.53
CV(%)	1.692	0.878	0.25	0.36	1.0	1.878
New Cal. Factor (%)	103.41	97.84	94.78	100.65	102.70	
Old Cal. Factors (%)	104.39	94.38	94.65	100.23	98.26	
Old TC1 (%)						99.11
New TC1 (%)						102.61

For Rapid Diagnostic Pvt. Ltd.



K. SRIKANTH
Service Engineer



N. Mahender Reddy
Zonal Service Manager



Calibration Certificate

Customer Name	Telangana Diagnostic Hub KOTHAGUEDEM
Address	Govt. Hospital, KHOTHAGUEDEM-507101
Contact Person	SUREKHA
Contact No.	6281135383
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-02000920
Instrument Status	Under Warranty
Calibration Done On	29.07.2022
Next Calibration Due On	28.07.2023

For Rapid Diagnostic Pvt. Ltd.



K. SRIKANTH
Service Engineer



N. Mahender Reddy
Zonal Service Manager