



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - FEBRUARY 2023



PC-1024

Lab Name **ANTLIA LAB N DIAGNOSTICS PVT LTD** Lab No **15833**
 Constituent Group **Chemistry I** Date of Result Entered : **21/02/2023**
 PT Item **Lyophilized human serum based** Date of Report Published : **08/03/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	1075	123.23	8.16	10.05	111 mg/dL	-1.22	0.61
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	648	129.71	8.40	10.90	104 mg/dL	-2.36	0.86
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	795	4.76	9.89	0.47	3.95 mg/dL	-1.72	0.03
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	821	1.58	18.29	0.29	1.5 mg/dL	-0.28	0.02
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	1100	4.63	9.80	0.45	4.6 g/dL	-0.07	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	730	2.91	8.69	0.25	3.1 g/dL	0.75	0.02
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	896	10.41	6.97	0.72	10 mg/dL	-0.57	0.05
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Any Analyser (Automation / Semi Automation)	390	4.03	12.07	0.49	4.4 mg/dL	0.76	0.05
9	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	1011	6.92	10.14	0.70	7.8 mg/dL	1.26	0.04
10	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	1046	98.66	7.43	7.33	95 mg/dL	-0.50	0.45
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	988	87.48	8.82	7.72	62 mg/dL	-3.3	0.49
12	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	714	28.16	19.26	5.42	31 mg/dL	0.52	0.41
13	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1147	120.46	4.12	4.96	102 mmol/L	-3.72	0.29
14	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1129	5.00	5.46	0.27	5.61 mmol/L	2.23	0.02
15	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	921	92.85	4.47	4.15	89 mmol/L	-0.93	0.27
16	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1021	190.14	15.11	28.72	202 U/L	0.41	1.80
17	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1025	49.30	15.70	7.74	50 U/L	0.09	0.48
18	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	733	178.51	26.55	47.40	191 U/L	0.26	3.50

19	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	449	79.72	17.30	13.80	75 U/L	-0.34	1.30
20	IRON	Ferrene (No Protein Removal)	Any Analyser (Automation / Semi Automation)	119	95.69	10.53	10.07	470 ug/dL	37.17	1.85
21	MAGNESIUM	Xylidyl Blue	Any Analyser (Automation / Semi Automation)	148	3.44	16.91	0.58	4 mg/dL	0.96	0.10

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

ANTLIA LAB N DIAGNOSTICS PVT LTD
GRANT ESTATE ROAD, BACKSIDE OF CIRCUIT HOSUE,
DUMKA
JHARKHAND814101

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****